

## Viscometer 3-1000 Low Temperature by





## **General Features**

In order to study gas viscosity under low temperature, ST designed the Viscometer 3-1000 LT. The gas viscometer is refrigerated in an oil bath to avoid complex mechanical assembly.

Specifications		
<b>Working Pressure</b>	1000 bar (15 000 psi)	
Chamber Volume	3 cc	
Temperature Range	-40 to 200 °C	
Temperature Stability	0.05°C	
Range Viscosity	0.01 MPa to 10 Pa	
Flow	0.01 to 40 cc/h	
Delta P Accuracy	0.0001 MPa	

The bath is equipped with cooling compressor, heating resistance and circulating pump.





## Viscometer 3-1000 Low Temperature by




To get quickly a stable Delta P, the chamber volume needs to be as small as possible due to the compressibility of the gas. In our case we need 3 min to stabilize the sample in a chamber volume of 3 ml. Two types of gaskets need to be used: silicone gaskets for  $-40^{\circ}$ C to  $+100^{\circ}$ C, and Viton gaskets for  $0^{\circ}$ C to  $+200^{\circ}$ C.


The viscometer is equipped with three pneumatic high pressure valves. It's also equipped with two pressure sensors with stability better than 1 mbar (10-4 MPa).

Capillary Tube Viscometer	
0.01 to 2 cp	Inside Ø 0.04 mm
0.1 to 20 cp	Inside Ø 0.08 mm
2 to 300 cp	Inside Ø 0.18 mm
80 to 10 000 cp	Inside Ø 0.5 mm









**High Position** 

